



REVISIONS					
REV	ECO	DESCRIPTIONS	DWN	CKD	APVD
1	✓	ORIGINATED	JNR		
			10/30/2001		



NOTES:

1. SS81B-0-0-0-0-0-0
2. PROCESS CONNS: 1/4" TUBE FITTINGS
3. INCHES (MM)
4. REF. DWG. 732-973BM

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE		 PANAMETRICS WALTHAM, MASS. 02154-3497	
FRACTIONS ±	DECIMALS .XX ± .01 .XXX ± .005	ANGLES ±	TITLE SAMPLE SYSTEM ASSEMBLY
DRAWN JNR 10/30/2001		ENGINEER	SIZE C
CHECKED			DRAWING NUMBER 732-973AS
RN			REV 1
MODEL NO. 973		SCALE	DO NOT SCALE DWG
REV		SHEET 1 OF 1	

PANAMETRICS
MOISTURE ANALYZER
SAMPLE SYSTEM

WARNING
BEFORE PRESSURIZING
ENSURE THAT THE PROBE
IS SECURELY INSTALLED
IN THE SAMPLE CELL

SAMPLE CELL

SAMPLE INLET

SAMPLE OUTLET

PANAMETRICS
101 CREEK
PHILADELPHIA, PA U.S.A.
TEL: 761-890-2716

SAMPLE SYSTEM
SERIAL NUMBER: 10042

PANAMETRICS
OXYGEN ANALYZER
SAMPLE SYSTEM

SET 0.0 TO 1.0 PSIG
SET 1.5 TO 2.0 SCFH

